


BP550 UV + NIR BLOCK/VISIBLE PASS



BP550 UV+NIR BLOCK/VISIBLE PASS

25.01.2006

Typical Data Points

Wavelength (nm)	Transmission (%)
1100.00	1.61
1090.00	1.44
1080.00	0.76
1070.00	0.37
1060.00	0.20
1050.00	0.13
1040.00	0.11
1030.00	0.10
1020.00	0.11
1010.00	0.16
1000.00	0.30
990.00	0.98
980.00	2.55
970.00	0.85
960.00	0.40
950.00	0.32
940.00	0.42
930.00	0.96
920.00	3.35
910.00	1.65
900.00	0.65
890.00	0.51
880.00	0.67
870.00	1.23
860.00	1.42
850.00	0.78
840.00	0.58
830.00	0.65
820.00	1.07
810.00	1.97

Wavelength (nm)	Transmission (%)
800.00	1.82
790.00	1.58
780.00	2.26
770.00	5.19
760.00	9.85
750.00	11.61
740.00	20.16
730.00	54.70
720.00	89.80
710.00	89.12
700.00	80.57
690.00	83.22
680.00	94.81
670.00	96.25
660.00	92.95
650.00	94.50
640.00	98.03
630.00	94.30
620.00	97.71
610.00	98.08
600.00	97.91
590.00	97.07
580.00	97.12
570.00	96.41
560.00	97.35
550.00	98.37
540.00	93.80
530.00	98.24
520.00	96.80
510.00	95.90

Wavelength (nm)	Transmission (%)
500.00	96.63
490.00	93.44
480.00	95.92
470.00	96.23
460.00	95.74
450.00	93.97
440.00	86.09
430.00	92.23
420.00	91.32
410.00	81.43
400.00	4.65
390.00	0.00
380.00	0.00
370.00	0.00
360.00	0.00
350.00	0.00
340.00	0.00
330.00	0.00
320.00	0.00
310.00	0.00
300.00	0.00
290.00	0.00
280.00	0.00
270.00	0.00
260.00	0.00
250.00	0.00
240.00	0.00
230.00	0.00
220.00	0.00
210.00	0.00
200.00	0.00